

STAR Post Shower Detector Review

The ESRC met at 10:00 AM in the MCR Conference Room to review the STAR Post Shower Detector.

Akio Ogawa made a presentation.

There were no significant points of concern, and two Action Items.

- 1) Have the electronics boards reviews by an Electrical Engineer qualified to provide EEI approval to each type of board. A sample of each type of board shall be marked with the EEI sticker. All subsequent boards of that type will also receive a sticker.
- 2) For the structural support panels of the Post Shower Detectors, provide to the Chair of the ESRC the MSDS for the G-10 used in construction.